

<b>Nominal (Nominal)</b>	<b>Patrón (Standard)</b>	<b>UBP (UUT)</b>	<b>Error (Error)</b>	<b>Exactitud (Accuracy)</b>	<b>Incertidumbre (Uncertainty)</b>
15 °C	15.121 °C	15.0 °C	-0.121 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	30.413 °C	30.2 °C	-0.213 °C	± 0.5 °C	± 4.0E-01 °C
35 °C	35.477 °C	35.2 °C	-0.277 °C	± 0.5 °C	± 4.0E-01 °C
40 %HR	40.591 %HR	41.3 %HR	0.709 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.325 %HR	50.2 %HR	-0.125 %HR	± 3.0 %HR	± 1.9E+00 %HR
90 %HR	90.386 %HR	90.6 %HR	0.214 %HR	± 3.0 %HR	± 1.9E+00 %HR