

| <b>Nominal (Nominal)</b> | <b>Patrón (Standard)</b> | <b>UBP (UUT)</b> | <b>Error (Error)</b> | <b>Exactitud (Accuracy)</b> | <b>Incertidumbre (Uncertainty)</b> |
|--------------------------|--------------------------|------------------|----------------------|-----------------------------|------------------------------------|
| 30 °C                    | 30.463 °C                | 30.3 °C          | -0.163 °C            | ± 0.5 °C                    | ± 4.0E-01 °C                       |
| 30 %HR                   | 30.263 %HR               | 33.1 %HR         | 2.837 %HR            | ± 3.0 %HR                   | ± 1.9E+00 %HR                      |
| 50 %HR                   | 50.049 %HR               | 51.9 %HR         | 1.851 %HR            | ± 3.0 %HR                   | ± 1.9E+00 %HR                      |
| 70 %HR                   | 70.243 %HR               | 69.1 %HR         | -1.143 %HR           | ± 3.0 %HR                   | ± 1.9E+00 %HR                      |