

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	30.463 °C	30.3 °C	-0.163 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.263 %HR	33.1 %HR	2.837 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.049 %HR	51.9 %HR	1.851 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.243 %HR	68.8 %HR	-1.443 %HR	± 3.0 %HR	± 1.9E+00 %HR