

Certificado No. (Certificate #):

67332

Equipo (Instrument): Reloj Indicador

Fecha de Calibración (Calibration Date):

2026-03-19

Marca (Brand): Mitutoyo

Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
1.27 mm	1.27 mm	0 µm	± 20 µm	± 5.8E+00 µm
2.54 mm	2.54 mm	0 µm	± 20 µm	± 5.8E+00 µm
3.30 mm	3.31 mm	5 µm	± 20 µm	± 5.8E+00 µm
4.57 mm	4.58 mm	5 µm	± 20 µm	± 5.8E+00 µm
5.08 mm	5.09 mm	5 µm	± 20 µm	± 5.8E+00 µm
7.62 mm	7.63 mm	5 µm	± 20 µm	± 5.8E+00 µm
10.16 mm	10.17 mm	5 µm	± 20 µm	± 5.8E+00 µm
12.70 mm	12.71 mm	5 µm	± 20 µm	± 5.8E+00 µm
17.78 mm	17.79 mm	10 µm	± 20 µm	± 5.8E+00 µm