

Certificado No. (Certificate #): 67333

Equipo (Instrument): Micrómetro de Exterior

Fecha de Calibración (Calibration Date): 2026-03-19

Marca (Brand): Mitutoyo

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)
50.8000 mm	50.800 mm	0 μm	$\pm 5 \mu\text{m}$	$\pm 5.8\text{E}+00 \mu\text{m}$
52.0700 mm	52.070 mm	0 μm	$\pm 5 \mu\text{m}$	$\pm 5.8\text{E}+00 \mu\text{m}$
53.8500 mm	53.850 mm	0 μm	$\pm 5 \mu\text{m}$	$\pm 5.8\text{E}+00 \mu\text{m}$
55.1200 mm	55.120 mm	0 μm	$\pm 5 \mu\text{m}$	$\pm 5.8\text{E}+00 \mu\text{m}$
58.4200 mm	58.420 mm	0 μm	$\pm 5 \mu\text{m}$	$\pm 5.8\text{E}+00 \mu\text{m}$
60.9600 mm	60.960 mm	0 μm	$\pm 5 \mu\text{m}$	$\pm 5.8\text{E}+00 \mu\text{m}$
76.2000 mm	76.200 mm	0 μm	$\pm 5 \mu\text{m}$	$\pm 5.8\text{E}+00 \mu\text{m}$
100.0000 mm	100.000 mm	0 μm	$\pm 5 \mu\text{m}$	$\pm 5.8\text{E}+00 \mu\text{m}$
101.6000 mm	101.600 mm	0 μm	$\pm 6 \mu\text{m}$	$\pm 5.8\text{E}+00 \mu\text{m}$
150.8000 mm	150.800 mm	0 μm	$\pm 7 \mu\text{m}$	$\pm 5.8\text{E}+00 \mu\text{m}$
250.8000 mm	250.800 mm	0 μm	$\pm 9 \mu\text{m}$	$\pm 5.8\text{E}+00 \mu\text{m}$