

Clase de Exactitud (Accuracy Class): Clase II (Class II)

Valor UBP	Valor Patrón	Error	EMP	Incertidumbre	Valor UBP	Valor Patrón	Error	EMP	Incertidumbre
m	m	mm	mm	(Uncertainty)	m	m	mm	mm	(Uncertainty)
0 a 0.3	0.29936	0.640	± 0.7	± 4.3E-01 mm					
0 a 0.6	0.59940	0.600	± 0.8	± 4.3E-01 mm					
0 a 0.9	0.89937	0.630	± 0.9	± 4.3E-01 mm					
0 a 1.2	1.19955	0.450	± 1	± 4.8E-01 mm					
0 a 1.5	1.49957	0.430	± 1.2	± 4.8E-01 mm					
0 a 1.8	1.79943	0.570	± 1.3	± 4.8E-01 mm					
0 a 2.1	2.09946	0.540	± 1.4	± 4.8E-01 mm					
0 a 2.4	2.39956	0.440	± 1.5	± 4.8E-01 mm					
0 a 2.7	2.69941	0.590	± 1.6	± 4.8E-01 mm					
0 a 3	2.99966	0.340	± 1.8	± 4.8E-01 mm					
