

<b>Patrón (Standard)</b>	<b>UBP (UUT)</b>	<b>Error (Error)</b>	<b>Exactitud (Accuracy)</b>	<b>Incertidumbre (Uncertainty)</b>
1799.888 s	1800 s	0.112 s	0.006%	± 0.01 % ± 1.7E+01 s
1799.974 s	1800 s	0.026 s	0.001%	± 0.01 % ± 1.7E+01 s
1800.102 s	1800 s	-0.102 s	-0.006%	± 0.01 % ± 1.7E+01 s