

<b>Patrón (Standard)</b>	<b>UBP (UUT)</b>	<b>Error (Error)</b>		<b>Exactitud (Accuracy)</b>	<b>Incertidumbre (Uncertainty)</b>	
1800.064 s	1800 s	-0.064 s	-0.004%	± 0.01 %	±	1.7E+01 s
1799.819 s	1800 s	0.181 s	0.010%	± 0.01 %	±	1.7E+01 s
1799.809 s	1800 s	0.191 s	0.011%	± 0.01 %	±	1.7E+01 s