

<b>Patrón (Standard)</b>	<b>UBP (UUT)</b>	<b>Error (Error)</b>	<b>Exactitud (Accuracy)</b>	<b>Incertidumbre (Uncertainty)</b>	
49.80 mmHg	50 mmHg	0.20 mmHg	± 3 mmHg	±	6.6E-01 mmHg
99.70 mmHg	100 mmHg	0.30 mmHg	± 3 mmHg	±	6.5E-01 mmHg
149.80 mmHg	150 mmHg	0.20 mmHg	± 3 mmHg	±	6.6E-01 mmHg
199.70 mmHg	200 mmHg	0.30 mmHg	± 3 mmHg	±	6.8E-01 mmHg
249.80 mmHg	250 mmHg	0.20 mmHg	± 3 mmHg	±	7.5E-01 mmHg
299.80 mmHg	300 mmHg	0.20 mmHg	± 3 mmHg	±	7.5E-01 mmHg