

| <b>Nominal</b> ( <i>Nominal</i> ) | <b>Patrón</b> ( <i>Standard</i> ) | <b>UBP</b> ( <i>UUT</i> ) | <b>Error</b> ( <i>Error</i> ) | <b>Exactitud</b> ( <i>Accuracy</i> ) | <b>Incertidumbre</b> ( <i>Uncertainty</i> ) |
|-----------------------------------|-----------------------------------|---------------------------|-------------------------------|--------------------------------------|---|
| 30 °C                             | 30.463 °C                         | 30.4 °C                   | -0.063 °C                     | ± 0.5 °C                             | ± 4.0E-01 °C                                |
| 30 %HR                            | 30.263 %HR                        | 31.7 %HR                  | 1.437 %HR                     | ± 3.0 %HR                            | ± 1.9E+00 %HR                               |
| 50 %HR                            | 50.049 %HR                        | 51.2 %HR                  | 1.151 %HR                     | ± 3.0 %HR                            | ± 1.9E+00 %HR                               |
| 70 %HR                            | 70.243 %HR                        | 69.0 %HR                  | -1.243 %HR                    | ± 3.0 %HR                            | ± 1.9E+00 %HR                               |