

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (M.P.E)	Incertidumbre (Uncertainty)	
5.01231 ml	5 ml	-0.01231 ml	± 0.03 ml	±	3.0E-01 µl
10.02323 ml	10 ml	-0.02323 ml	± 0.06 ml	±	5.9E-01 µl
25.05977 ml	25 ml	-0.05977 ml	± 0.15 ml	±	1.5E+00 µl