

<b>Patrón (Standard)</b>	<b>UBP (UUT)</b>	<b>Error (Error)</b>	<b>E.M.P (Tolerance)</b>	<b>Incertidumbre (Uncertainty)</b>
52.65 $\mu\text{m}$	52.98 $\mu\text{m}$	0.33 $\mu\text{m}$	$\pm 1.6 \mu\text{m}$	$\pm 5.8\text{E-}02 \mu\text{m}$
155.23 $\mu\text{m}$	155.0 $\mu\text{m}$	-0.23 $\mu\text{m}$	$\pm 4.7 \mu\text{m}$	$\pm 6.2\text{E-}02 \mu\text{m}$
257.12 $\mu\text{m}$	257.0 $\mu\text{m}$	-0.12 $\mu\text{m}$	$\pm 7.7 \mu\text{m}$	$\pm 7.0\text{E-}02 \mu\text{m}$
511.72 $\mu\text{m}$	511.4 $\mu\text{m}$	-0.32 $\mu\text{m}$	$\pm 15.4 \mu\text{m}$	$\pm 9.7\text{E-}02 \mu\text{m}$