

Certificado No. (Certificate #):

67385

Equipo (Instrument): Calibrador Pie de Rey

Fecha de Calibración (Calibration Date):

2026-03-23

Marca (Brand):

TOTAL

Tipo de Medida (Measurement Type)	Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)
Exterior / (Exterior)	10.16 mm	10.16 mm	0 $\mu\text{m}$	$\pm 30 \mu\text{m}$	$\pm 7.4\text{E}+00 \mu\text{m}$
Exterior / (Exterior)	25.40 mm	25.41 mm	10 $\mu\text{m}$	$\pm 30 \mu\text{m}$	$\pm 7.4\text{E}+00 \mu\text{m}$
Exterior / (Exterior)	100.00 mm	100.00 mm	0 $\mu\text{m}$	$\pm 30 \mu\text{m}$	$\pm 7.4\text{E}+00 \mu\text{m}$
Exterior / (Exterior)	131.50 mm	131.50 mm	0 $\mu\text{m}$	$\pm 30 \mu\text{m}$	$\pm 7.4\text{E}+00 \mu\text{m}$
Exterior / (Exterior)	148.26 mm	148.26 mm	0 $\mu\text{m}$	$\pm 30 \mu\text{m}$	$\pm 7.4\text{E}+00 \mu\text{m}$
Interior / (Interior)	76.20 mm	76.20 mm	0 $\mu\text{m}$	$\pm 30 \mu\text{m}$	$\pm 8.2\text{E}+00 \mu\text{m}$
Interior / (Interior)	100.00 mm	100.00 mm	0 $\mu\text{m}$	$\pm 30 \mu\text{m}$	$\pm 8.2\text{E}+00 \mu\text{m}$
Profundidad / (Depth)	50.80 mm	50.80 mm	0 $\mu\text{m}$	$\pm 30 \mu\text{m}$	$\pm 7.1\text{E}+00 \mu\text{m}$
Profundidad / (Depth)	100.00 mm	100.00 mm	0 $\mu\text{m}$	$\pm 30 \mu\text{m}$	$\pm 7.1\text{E}+00 \mu\text{m}$