

Certificado No. (Certificate #): 67394

Equipo (Instrument): Micrómetro de Exterior

Fecha de Calibración (Calibration Date): 2026-03-19

Marca (Brand): Mitutoyo

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)
150.80 mm	150.80 mm	0 µm	± 7 µm	± 5.8E+00 µm
152.07 mm	152.07 mm	0 µm	± 7 µm	± 5.8E+00 µm
153.85 mm	153.85 mm	0 µm	± 7 µm	± 5.8E+00 µm
155.12 mm	155.12 mm	0 µm	± 7 µm	± 5.8E+00 µm
158.42 mm	158.42 mm	0 µm	± 7 µm	± 5.8E+00 µm
160.96 mm	160.96 mm	0 µm	± 7 µm	± 5.8E+00 µm
164.77 mm	164.77 mm	0 µm	± 7 µm	± 5.8E+00 µm
166.80 mm	166.80 mm	0 µm	± 7 µm	± 5.8E+00 µm
168.33 mm	168.33 mm	0 µm	± 7 µm	± 5.8E+00 µm
171.12 mm	171.12 mm	0 µm	± 7 µm	± 5.8E+00 µm
173.66 mm	173.66 mm	0 µm	± 7 µm	± 5.8E+00 µm