

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	30.283 °C	30.5 °C	0.217 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	29.537 %HR	31.4 %HR	1.863 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.631 %HR	53.3 %HR	2.669 %HR	± 3.0 %HR	± 1.9E+00 %HR
75 %HR	75.243 %HR	74.5 %HR	-0.743 %HR	± 3.0 %HR	± 1.9E+00 %HR