

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	30.283 °C	30.5 °C	0.217 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	29.537 %HR	32.0 %HR	2.463 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.631 %HR	52.8 %HR	2.169 %HR	± 3.0 %HR	± 1.9E+00 %HR
75 %HR	75.243 %HR	73.8 %HR	-1.443 %HR	± 3.0 %HR	± 1.9E+00 %HR