

| Nominal (Nominal) | Patrón (Standard) | UBP (UUT) | Error (Error) | Exactitud (Accuracy) | Incertidumbre (Uncertainty) |
|--------------------------|--------------------------|------------------|----------------------|-----------------------------|------------------------------------|
| 30 °C | 30.283 °C | 30.4 °C | 0.117 °C | ± 0.5 °C | ± 4.0E-01 °C |
| 30 %HR | 29.537 %HR | 32.3 %HR | 2.763 %HR | ± 3.0 %HR | ± 1.9E+00 %HR |
| 50 %HR | 50.631 %HR | 52.1 %HR | 1.469 %HR | ± 3.0 %HR | ± 1.9E+00 %HR |
| 75 %HR | 75.243 %HR | 74.3 %HR | -0.943 %HR | ± 3.0 %HR | ± 1.9E+00 %HR |