

<b>Nominal</b> ( <i>Nominal</i> )	<b>Patrón</b> ( <i>Standard</i> )	<b>UBP</b> ( <i>UUT</i> )	<b>Error</b> ( <i>Error</i> )	<b>Exactitud</b> ( <i>Accuracy</i> )	<b>Incertidumbre</b> ( <i>Uncertainty</i> )
30 °C	30.283 °C	30.4 °C	0.117 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	29.537 %HR	32.2 %HR	2.663 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.631 %HR	52.2 %HR	1.569 %HR	± 3.0 %HR	± 1.9E+00 %HR
75 %HR	75.243 %HR	74.6 %HR	-0.643 %HR	± 3.0 %HR	± 1.9E+00 %HR