

<b>Nominal (Nominal)</b>	<b>Patrón (Standard)</b>	<b>UBP (UUT)</b>	<b>Error (Error)</b>	<b>Exactitud (Accuracy)</b>	<b>Incertidumbre (Uncertainty)</b>
30 °C	30.614 °C	29.5 °C	-1.114 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	30.570 %HR	31.0 %HR	0.430 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	50.120 %HR	51.0 %HR	0.880 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	70.570 %HR	68.0 %HR	-2.570 %HR	± 5.0 %HR	± 1.9E+00 %HR