

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
-25 °C	-25.240 °C	-24.7 °C	0.540 °C	± 1.1 °C	± 4.0E-01 °C
0 °C	-0.041 °C	0.3 °C	0.341 °C	± 0.5 °C	± 4.0E-01 °C
5 °C	4.973 °C	5.3 °C	0.327 °C	± 0.5 °C	± 4.0E-01 °C
10 °C	10.158 °C	10.2 °C	0.042 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.069 °C	24.8 °C	-0.269 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	30.072 °C	29.8 °C	-0.272 °C	± 0.5 °C	± 4.0E-01 °C
40 °C	40.301 °C	40.1 °C	-0.201 °C	± 0.5 °C	± 4.0E-01 °C

