

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	30.463 °C	30.4 °C	-0.063 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	30.263 %HR	30.0 %HR	-0.263 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	50.049 %HR	50.0 %HR	-0.049 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	70.243 %HR	70.0 %HR	-0.243 %HR	± 5.0 %HR	± 1.9E+00 %HR