

<b>Nominal</b> ( <i>Nominal</i> )	<b>Patrón</b> ( <i>Standard</i> )	<b>UBP</b> ( <i>UUT</i> )	<b>Error</b> ( <i>Error</i> )	<b>Exactitud</b> ( <i>Accuracy</i> )	<b>Incertidumbre</b> ( <i>Uncertainty</i> )
30 °C	30.463 °C	30.5 °C	0.037 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	30.263 %HR	27.0 %HR	-3.263 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	50.049 %HR	48.0 %HR	-2.049 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	70.243 %HR	69.0 %HR	-1.243 %HR	± 5.0 %HR	± 1.9E+00 %HR