

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
0 °C	-0.572 °C	-0.3 °C	0.272 °C	± 1.0 °C	± 4.0E-01 °C
20 °C	20.195 °C	20.3 °C	0.105 °C	± 1.0 °C	± 4.0E-01 °C
35 °C	35.517 °C	35.4 °C	-0.117 °C	± 1.0 °C	± 4.0E-01 °C
# 10 %HR	10.539 %HR	15.0 %HR	4.461 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	49.711 %HR	52.0 %HR	2.289 %HR	± 5.0 %HR	± 1.9E+00 %HR
80 %HR	81.501 %HR	77.0 %HR	-4.501 %HR	± 5.0 %HR	± 1.9E+00 %HR