

| Nominal (Nominal) | Patrón (Standard) | UBP (UUT) | Error (Error) | Exactitud (Accuracy) | Incertidumbre (Uncertainty) |
|--------------------------|--------------------------|------------------|----------------------|-----------------------------|------------------------------------|
| 0 °C | -0.572 °C | -0.3 °C | 0.272 °C | ± 1.0 °C | ± 4.0E-01 °C |
| 20 °C | 20.195 °C | 20.3 °C | 0.105 °C | ± 1.0 °C | ± 4.0E-01 °C |
| 35 °C | 35.517 °C | 35.3 °C | -0.217 °C | ± 1.0 °C | ± 4.0E-01 °C |
| # 10 %HR | 10.539 %HR | 15.0 %HR | 4.461 %HR | ± 5.0 %HR | ± 1.9E+00 %HR |
| 50 %HR | 49.711 %HR | 52.0 %HR | 2.289 %HR | ± 5.0 %HR | ± 1.9E+00 %HR |
| 80 %HR | 81.501 %HR | 79.0 %HR | -2.501 %HR | ± 5.0 %HR | ± 1.9E+00 %HR |