

Certificado No. (Certificate #):

67477

Equipo (Instrument): Nivel Digital

Fecha de Calibración (Calibration Date):

2026-02-23

Marca (Brand):

BOSCH

Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
5 °	4.9 °	-0.1 °	± 0.2 °	± 5.80E-02 °
15 °	14.9 °	-0.1 °	± 0.2 °	± 5.80E-02 °
25 °	25.0 °	0.0 °	± 0.2 °	± 5.80E-02 °
30 °	30.0 °	0.0 °	± 0.2 °	± 5.80E-02 °
45 °	45.0 °	0.0 °	± 0.2 °	± 5.80E-02 °
50 °	50.0 °	0.0 °	± 0.2 °	± 5.80E-02 °
60 °	59.9 °	-0.1 °	± 0.2 °	± 5.80E-02 °
75 °	75.0 °	0.0 °	± 0.2 °	± 5.80E-02 °
90 °	90.0 °	0.0 °	± 0.2 °	± 5.80E-02 °