

| Nominal (Nominal) | Patrón (Standard) | UBP (UUT) | Error (Error) | Exactitud (Accuracy) | Incertidumbre (Uncertainty) |
|-------------------|-------------------|-----------|---------------|----------------------|-----------------------------|
| 10 °C | 10.519 °C | 10.5 °C | -0.019 °C | ± 0.5 °C | ± 4.0E-01 °C |
| 25 °C | 25.420 °C | 25.3 °C | -0.120 °C | ± 0.5 °C | ± 4.0E-01 °C |
| 30 °C | 30.614 °C | 30.6 °C | -0.014 °C | ± 0.5 °C | ± 4.0E-01 °C |
| 40 °C | 39.980 °C | 39.8 °C | -0.180 °C | ± 0.5 °C | ± 4.0E-01 °C |

