

| Nominal (Nominal) | Patrón (Standard) | UBP (UUT) | Error (Error) | Exactitud (Accuracy) | Incertidumbre (Uncertainty) |
|-------------------|-------------------|-----------|---------------|----------------------|-----------------------------|
| 10 °C | 10.519 °C | 10.8 °C | 0.281 °C | ± 0.5 °C | ± 4.0E-01 °C |
| 25 °C | 25.420 °C | 25.2 °C | -0.220 °C | ± 0.5 °C | ± 4.0E-01 °C |
| 30 °C | 30.614 °C | 30.3 °C | -0.314 °C | ± 0.5 °C | ± 4.0E-01 °C |
| 40 °C | 39.980 °C | 40.3 °C | 0.320 °C | ± 0.5 °C | ± 4.0E-01 °C |

