

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 %HR	30.570 %HR	29.2 %HR	-1.370 %HR	± 5 %HR	± 1.9E+00 %HR
50 %HR	50.120 %HR	48.4 %HR	-1.720 %HR	± 5 %HR	± 1.9E+00 %HR
70 %HR	70.263 %HR	67.3 %HR	-2.963 %HR	± 5 %HR	± 1.9E+00 %HR
90 %HR	90.080 %HR	86.2 %HR	-3.880 %HR	± 5 %HR	± 1.9E+00 %HR

