

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
10 °C	10.519 °C	10.1 °C	-0.419 °C	± 0.5 °C	± 4.0E-01 °C
25 °C	25.420 °C	25.7 °C	0.280 °C	± 0.5 °C	± 4.0E-01 °C
30 °C	30.614 °C	30.6 °C	-0.014 °C	± 0.5 °C	± 4.0E-01 °C
40 °C	39.980 °C	40.4 °C	0.420 °C	± 0.5 °C	± 4.0E-01 °C

