

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 %HR	30.570 %HR	29.9 %HR	-0.670 %HR	± 5 %HR	± 1.9E+00 %HR
50 %HR	50.120 %HR	49.9 %HR	-0.220 %HR	± 5 %HR	± 1.9E+00 %HR
70 %HR	70.263 %HR	69.7 %HR	-0.563 %HR	± 5 %HR	± 1.9E+00 %HR
90 %HR	90.080 %HR	86.5 %HR	-3.580 %HR	± 5 %HR	± 1.9E+00 %HR

