

<b>Nominal (Nominal)</b>	<b>Patrón (Standard)</b>	<b>UBP (UUT)</b>	<b>Error (Error)</b>	<b>Exactitud (Accuracy)</b>	<b>Incertidumbre (Uncertainty)</b>
30 °C	30.514 °C	30.4 °C	-0.114 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	29.834 %HR	32.7 %HR	2.866 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	49.865 %HR	52.5 %HR	2.635 %HR	± 3.0 %HR	± 1.9E+00 %HR
75 %HR	75.938 %HR	75.5 %HR	-0.438 %HR	± 3.0 %HR	± 1.9E+00 %HR