

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	30.514 °C	30.4 °C	-0.114 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	29.834 %HR	32.4 %HR	2.566 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	49.865 %HR	52.3 %HR	2.435 %HR	± 3.0 %HR	± 1.9E+00 %HR
75 %HR	75.938 %HR	75.7 %HR	-0.238 %HR	± 3.0 %HR	± 1.9E+00 %HR