

<b>Nominal (Nominal)</b>	<b>Patrón (Standard)</b>	<b>UBP (UUT)</b>	<b>Error (Error)</b>	<b>Exactitud (Accuracy)</b>	<b>Incertidumbre (Uncertainty)</b>
30 °C	30.363 °C	30.5 °C	0.137 °C	± 0.5 °C	± 4.0E-01 °C
35 %HR	35.580 %HR	36.1 %HR	0.520 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.120 %HR	52.6 %HR	2.480 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.243 %HR	69.5 %HR	-0.743 %HR	± 3.0 %HR	± 1.9E+00 %HR