

| Nominal (<i>Nominal</i>) | Patrón (<i>Standard</i>) | UBP (<i>UUT</i>) | Error (<i>Error</i>) | Exactitud (<i>Accuracy</i>) | Incertidumbre (<i>Uncertainty</i>) |
|-----------------------------------|-----------------------------------|---------------------------|-------------------------------|--------------------------------------|---|
| 30 °C | 30.363 °C | 30.5 °C | 0.137 °C | ± 0.5 °C | ± 4.0E-01 °C |
| 35 %HR | 35.580 %HR | 34.2 %HR | -1.380 %HR | ± 3.0 %HR | ± 1.9E+00 %HR |
| 50 %HR | 50.120 %HR | 51.7 %HR | 1.580 %HR | ± 3.0 %HR | ± 1.9E+00 %HR |
| 70 %HR | 70.243 %HR | 70.3 %HR | 0.057 %HR | ± 3.0 %HR | ± 1.9E+00 %HR |