

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	30.363 °C	30.5 °C	0.137 °C	± 0.5 °C	± 4.0E-01 °C
35 %HR	35.580 %HR	34.5 %HR	-1.080 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.120 %HR	52.1 %HR	1.980 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.243 %HR	67.6 %HR	-2.643 %HR	± 3.0 %HR	± 1.9E+00 %HR