

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	30.403 °C	30.4 °C	-0.003 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.519 %HR	32.5 %HR	1.981 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.294 %HR	51.4 %HR	1.106 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.018 %HR	68.7 %HR	-1.318 %HR	± 3.0 %HR	± 1.9E+00 %HR