

<b>Nominal (Nominal)</b>	<b>Patrón (Standard)</b>	<b>UBP (UUT)</b>	<b>Error (Error)</b>	<b>Exactitud (Accuracy)</b>	<b>Incertidumbre (Uncertainty)</b>
30 °C	30.373 °C	30.3 °C	-0.073 °C	± 0.5 °C	± 4.0E-01 °C
30 %HR	30.519 %HR	33.2 %HR	2.681 %HR	± 3.0 %HR	± 1.9E+00 %HR
50 %HR	50.294 %HR	53.1 %HR	2.806 %HR	± 3.0 %HR	± 1.9E+00 %HR
70 %HR	70.018 %HR	69.1 %HR	-0.918 %HR	± 3.0 %HR	± 1.9E+00 %HR