

| Nominal (Nominal) | Patrón (Standard) | UBP (UUT) | Error (Error) | Exactitud (Accuracy) | Incertidumbre (Uncertainty) |
|--------------------------|--------------------------|------------------|----------------------|-----------------------------|------------------------------------|
| 30 °C | 30.373 °C | 30.3 °C | -0.073 °C | ± 0.5 °C | ± 4.0E-01 °C |
| 30 %HR | 30.519 %HR | 33.4 %HR | 2.881 %HR | ± 3.0 %HR | ± 1.9E+00 %HR |
| 50 %HR | 50.294 %HR | 51.4 %HR | 1.106 %HR | ± 3.0 %HR | ± 1.9E+00 %HR |
| 70 %HR | 70.018 %HR | 67.8 %HR | -2.218 %HR | ± 3.0 %HR | ± 1.9E+00 %HR |