

| Tipo de Medida (<i>Measurement Type</i>) | Patrón (<i>Standard</i>) | UBP (<i>UT</i>) | Error (<i>Error</i>) | E.M.P (<i>Tolerance</i>) | Incertidumbre (<i>Uncertainty</i>) |
|---|-----------------------------------|--------------------------|-------------------------------|-----------------------------------|---|
| Exterior / (Exterior) | 40 mm | 39.99 mm | -10 μm | $\pm 30 \mu\text{m}$ | $\pm 7.4\text{E}+00 \mu\text{m}$ |
| Exterior / (Exterior) | 80 mm | 79.99 mm | -10 μm | $\pm 30 \mu\text{m}$ | $\pm 7.6\text{E}+00 \mu\text{m}$ |
| Exterior / (Exterior) | 120 mm | 119.99 mm | -10 μm | $\pm 30 \mu\text{m}$ | $\pm 7.8\text{E}+00 \mu\text{m}$ |
| Exterior / (Exterior) | 160 mm | 159.99 mm | -10 μm | $\pm 30 \mu\text{m}$ | $\pm 8.1\text{E}+00 \mu\text{m}$ |
| Exterior / (Exterior) | 200 mm | 200.00 mm | 0 μm | $\pm 30 \mu\text{m}$ | $\pm 8.4\text{E}+00 \mu\text{m}$ |
| Interior / (Interior) | 20 mm | 20.01 mm | 10 μm | $\pm 30 \mu\text{m}$ | $\pm 8.2\text{E}+00 \mu\text{m}$ |
| Interior / (Interior) | 180 mm | 180.01 mm | 10 μm | $\pm 30 \mu\text{m}$ | $\pm 8.3\text{E}+00 \mu\text{m}$ |
| Profundidad / (Depth) | 20 mm | 20.01 mm | 10 μm | $\pm 30 \mu\text{m}$ | $\pm 7.1\text{E}+00 \mu\text{m}$ |
| Profundidad / (Depth) | 180 mm | 180.01 mm | 10 μm | $\pm 30 \mu\text{m}$ | $\pm 7.9\text{E}+00 \mu\text{m}$ |