

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)	
130.3 μm	130.6 μm	0.3 μm	$\pm 6.53 \mu\text{m}$	\pm	5.8E-01 μm
523.0 μm	523.0 μm	0.0 μm	$\pm 26.15 \mu\text{m}$	\pm	5.8E-01 μm