

Clase de Exactitud (Accuracy Class): Clase I (Class I)

Valor UBP	Valor Patrón	Error	EMP	Incertidumbre (Uncertainty)	Distancia de objetos (Colocados)	Distancia detectada		Error	Incertidumbre (Uncertainty)
m	m	mm	mm		m	m	mm	mm	
0 a 1	0.99965	0.350	± 0.4	± 4.3E-01 mm	1	1.00047		0.00047	4.3E-01 mm
0 a 2	2.00012	-0.120	± 0.6	± 4.8E-01 mm	2	2.00099		0.00099	4.8E-01 mm
0 a 3	2.99955	0.450	± 0.8	± 4.8E-01 mm	3	3.00026		0.00026	4.8E-01 mm
0 a 4	3.99990	0.100	± 1.0	± 4.8E-01 mm	4	4.00045		0.00045	4.8E-01 mm
0 a 5	4.99978	0.220	± 1.2	± 4.8E-01 mm	5	5.00021		0.00021	4.8E-01 mm
0 a 6	6.00022	-0.220	± 1.4	± 3.6E-01 mm	6	6.00053		0.00053	3.6E-01 mm
0 a 7	7.00003	-0.030	± 1.6	± 3.6E-01 mm	7	7.00041		0.00041	3.6E-01 mm
0 a 8	7.99983	0.170	± 1.8	± 3.6E-01 mm	8	8.00029		0.00028	3.6E-01 mm
0 a 9	9.00024	-0.240	± 2.0	± 3.6E-01 mm	9	9.00011		0.00011	3.6E-01 mm
0 a 10	9.99918	0.820	± 2.2	± 3.6E-01 mm	10	9.94590		-0.05410	3.6E-01 mm
0 a 11	11.00014	-0.140	± 2.4	± 3.6E-01 mm	11	11.00025		0.00024	3.6E-01 mm
0 a 12	12.00048	-0.480	± 2.6	± 3.6E-01 mm	12	12.00081		0.00081	3.6E-01 mm
0 a 13	13.00039	-0.390	± 2.8	± 3.6E-01 mm	13	13.00058		0.00058	3.6E-01 mm
0 a 14	14.00044	-0.440	± 3.0	± 3.6E-01 mm	14	14.00051		0.00051	3.6E-01 mm
0 a 15	15.00055	-0.550	± 3.2	± 3.6E-01 mm	15	15.00066		0.00066	3.6E-01 mm