

Clase de Exactitud (Accuracy Class): Clase I (Class I)

Valor UBP	Valor Patrón	Error	EMP	Incertidumbre	Valor UBP	Valor Patrón	Error	EMP	Incertidumbre
m	m	mm	mm	(Uncertainty)	m	m	mm	mm	(Uncertainty)
0 a 1	0.99966	0.340	± 0.4	± 4.3E-01 mm					
0 a 2	2.00015	-0.150	± 0.6	± 4.8E-01 mm					
0 a 3	2.99962	0.380	± 0.8	± 4.8E-01 mm					
0 a 4	3.99989	0.110	± 1.0	± 4.8E-01 mm					
0 a 5	4.99979	0.210	± 1.2	± 4.8E-01 mm					
0 a 6	6.00027	-0.270	± 1.4	± 3.6E-01 mm					
0 a 7	7.00008	-0.080	± 1.6	± 3.6E-01 mm					
0 a 8	7.99980	0.200	± 1.8	± 3.6E-01 mm					
0 a 9	9.00030	-0.300	± 2.0	± 3.6E-01 mm					
0 a 10	9.99913	0.870	± 2.2	± 3.6E-01 mm					
0 a 11	11.00016	-0.160	± 2.4	± 3.6E-01 mm					
0 a 12	12.00048	-0.480	± 2.6	± 3.6E-01 mm					
0 a 13	13.00037	-0.370	± 2.8	± 3.6E-01 mm					
0 a 14	14.00043	-0.430	± 3.0	± 3.6E-01 mm					
0 a 15	15.00053	-0.530	± 3.2	± 3.6E-01 mm					
