

Nominal (Nominal)	Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)
30 °C	30.123 °C	29.8 °C	-0.323 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	30.417 %HR	31 %HR	0.583 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	49.558 %HR	51 %HR	1.442 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	70.478 %HR	68 %HR	-2.478 %HR	± 5.0 %HR	± 1.9E+00 %HR