

<b>Nominal (Nominal)</b>	<b>Patrón (Standard)</b>	<b>UBP (UUT)</b>	<b>Error (Error)</b>	<b>Exactitud (Accuracy)</b>	<b>Incertidumbre (Uncertainty)</b>
30 °C	30.123 °C	30.1 °C	-0.023 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	30.417 %HR	31 %HR	0.583 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	49.558 %HR	50 %HR	0.442 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	70.478 %HR	70 %HR	-0.478 %HR	± 5.0 %HR	± 1.9E+00 %HR