

| <b>Nominal (Nominal)</b> | <b>Patrón (Standard)</b> | <b>UBP (UUT)</b> | <b>Error (Error)</b> | <b>Exactitud (Accuracy)</b> | <b>Incertidumbre (Uncertainty)</b> |
|--------------------------|--------------------------|------------------|----------------------|-----------------------------|------------------------------------|
| 30 °C                    | 30.123 °C                | 30.0 °C          | -0.123 °C            | ± 1.0 °C                    | ± 4.0E-01 °C                       |
| 30 %HR                   | 30.417 %HR               | 31 %HR           | 0.583 %HR            | ± 5.0 %HR                   | ± 1.9E+00 %HR                      |
| 50 %HR                   | 49.558 %HR               | 50 %HR           | 0.442 %HR            | ± 5.0 %HR                   | ± 1.9E+00 %HR                      |
| 70 %HR                   | 70.478 %HR               | 70 %HR           | -0.478 %HR           | ± 5.0 %HR                   | ± 1.9E+00 %HR                      |