

Nominal (<i>Nominal</i>)	Patrón (<i>Standard</i>)	UBP (<i>UUT</i>)	Error (<i>Error</i>)	Exactitud (<i>Accuracy</i>)	Incertidumbre (<i>Uncertainty</i>)
30 °C	30.123 °C	30.3 °C	0.177 °C	± 1.0 °C	± 4.0E-01 °C
30 %HR	30.417 %HR	32 %HR	1.583 %HR	± 5.0 %HR	± 1.9E+00 %HR
50 %HR	49.558 %HR	50 %HR	0.442 %HR	± 5.0 %HR	± 1.9E+00 %HR
70 %HR	70.478 %HR	73 %HR	2.522 %HR	± 5.0 %HR	± 1.9E+00 %HR