

---

<b>Patrón (Standard)</b>	<b>UBP (UUT)</b>	<b>Error (Error)</b>	<b>Exactitud (Accuracy)</b>	<b>Incertidumbre (Uncertainty)</b>	
60 RPM	59.9 RPM	-0.1 RPM	± 0.03 RPM	±	5.8E-02 RPM
600 RPM	600.2 RPM	0.2 RPM	± 0.3 RPM	±	5.8E-02 RPM
6000 RPM	5999 RPM	-1.0 RPM	± 3 RPM	±	5.8E-01 RPM
60000 RPM	59997 RPM	-3.0 RPM	± 30 RPM	±	9.0E-01 RPM
99960 RPM	99957 RPM	-3.0 RPM	± 50 RPM	±	9.3E-01 RPM