
Patrón (Standard)	UBP (UUT)	Error (Error)	Exactitud (Accuracy)	Incertidumbre (Uncertainty)	
60 RPM	59.9 RPM	-0.1 RPM	± 0.03 RPM	±	5.8E-02 RPM
600 RPM	600.2 RPM	0.2 RPM	± 0.3 RPM	±	5.8E-02 RPM
6000 RPM	6000 RPM	0.0 RPM	± 3 RPM	±	5.8E-01 RPM
60000 RPM	59997 RPM	-3.0 RPM	± 30 RPM	±	9.0E-01 RPM
99960 RPM	99957 RPM	-3.0 RPM	± 50 RPM	±	9.3E-01 RPM