

Patrón (Standard)	UBP (UUT)	Error (Error)	E.M.P (Tolerance)	Incertidumbre (Uncertainty)	
23.2 μm	23.3 μm	0.1 μm	± 1.5 μm	±	2.1E-01 μm
51.1 μm	51.2 μm	0.1 μm	± 2.0 μm	±	2.1E-01 μm
125.1 μm	125 μm	-0.1 μm	± 3.5 μm	±	2.1E-01 μm
478.0 μm	479 μm	1.0 μm	± 6.0 μm	±	2.2E-01 μm
952 μm	953 μm	1.0 μm	± 11 μm	±	2.5E-01 μm