

<b>Patrón (Standard)</b>	<b>UBP (UUT)</b>	<b>Error (Error)</b>	<b>E.M.P (Tolerance)</b>	<b>Incertidumbre (Uncertainty)</b>	
3.0000 in	3.0000 in	0.0000 in	± 0.0002 in	±	#REF! in
3.0625 in	3.0625 in	0.0000 in	± 0.0002 in	±	#REF! in
3.1000 in	3.1000 in	0.0000 in	± 0.0002 in	±	#REF! in
3.1250 in	3.1250 in	0.0000 in	± 0.0002 in	±	#REF! in
3.2000 in	3.2000 in	0.0000 in	± 0.0002 in	±	#REF! in
3.2500 in	3.2500 in	0.0000 in	± 0.0002 in	±	#REF! in
3.3000 in	3.3000 in	0.0000 in	± 0.0002 in	±	#REF! in
3.4000 in	3.4000 in	0.0000 in	± 0.0002 in	±	#REF! in
3.5000 in	3.5000 in	0.0000 in	± 0.0002 in	±	#REF! in
3.8000 in	3.8000 in	0.0000 in	± 0.0002 in	±	5.8E-05 in
4.0000 in	4.0000 in	0.0000 in	± 0.0002 in	±	5.8E-05 in